

FEATURES



- 94dB And 114dB Sound Calibrator At 1KHz
- Accurate And Simple To Use
- Fits 1inch And 1/2inch Microphones
- Low Battery Indicator
- IEC 60942(2003) Class 2, Ans1.40-1984

SPECIFICATIONS

Output Sound Pressure Levels:
114dB and 94dB re 20uPa under reference Conditions
Output Frequency: 1000Hz \pm 2%
Total Harmonic Distortion: < 4%
Reference Condition:
Temperature 23°C(73°F)
Static press 103hpa
Relative humidity 50%
Accuracy of sound pressure level:
Understated reference environment conditions \pm 0.5dB
Battery test:
Operate: Green LED indicator
Low battery: Red LED indicator
Battery: 9V battery
Dimensions: 113x63x44 mm
Weight: Approx.170g



CENTER[®]
CENTER TECHNOLOGY CORP.



4F, No.415, Jung-Jeng Road, Shu-Lin Dist.,
New Taipei City 238, Taiwan
TEL: 886-2-26763926 E-Mail : center@centertek.com
FAX: 886-2-26763925 http ://www.centertek.com

Distributed by: